Attachment E

U.S. DEPARTMENT OF AGRICULTURE
FOOD SAFETY AND INSPECTION SERVICE
INTERNATIONAL PROGRAMS

FOREIGN COUNTRY LABORATORY REVIEW

REVIEW DATE: 6/26/01

NAME OF FOREIGN LABORATORY: Canadian Food Inspection Agency, Saskatoon Laboratory

FOREIGN GOV'T AGENCY: Canadian Food Insp. Agency

CITY & COUNTRY: Saskatoon, Canada

ADDRESS OF LABORATORY:

NAME OF REVIEWER: Dr. Oto Urban

NAME OF FOREIGN OFFICIAL: Dr. James D. MacNeil

	Residue Code/Name ▶			100	111	200	203	300	400	500	SPC	IVE			
	REVIEW ITEMS	ITEM #													
SAMPLING PROCEDURES	Sample Handling	01	EVALUATION CODE	A	O	A	A	A	O	A	O	A			
	Sampling Frequency	02		A	O	A	A	A	O	A	O	A			
	Timely Analyses	03		A	O	A	A	A	O	A	O	C			
	Compositing Procedure	04		O	O	O	O	O	O	O	O	O			
	Interpret Comp Data	05		O	O	O	O	O	O	O	O	O			
	Data Reporting	06		A	O	A	A	A	O	A	O	A			
ANALYTICAL PROCEDURES	Acceptable Method	07	EVALUATION CODE	A	O	A	A	A	O	A	O	A			
	Correct Tissue(s)	08		A	O	A	A	A	O	A	O	A			
	Equipment Operation	09		A	O	A	A	A	O	A	O	A			
	Instrument Printouts	10		A	O	A	A	A	O	A	O	A			
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11	EVALUATION CODE	A	O	A	A	A	O	A	O	A			
	Recovery Frequency	12		A	O	A	O	A	O	O	O	A			
	Percent Recovery	13		A	O	A	O	A	O	O	O	A			
	Check Sample Frequency	14		A	O	A	A	A	O	A	O	A			
	All analyst w/Check Samples	15		C	O	C	C	C	O	C	O	C			
	Corrective Actions	16		A	O	A	A	A	O	A	O	A			
	International Check Samples	17		A	O	A	A	A	O	A	O	A			
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL. CODE	A	O	A	A	A	O	A	O	A			
OTHER REVIEW		19	EVAL. CODE												
		20													

FSIS FORM 9520-4 (9/96)

FOREIGN COUNTRY LABORATORY REVIEW (Comment Sheet)	REVIEW DATE	NAME OF FOREIGN LABORATORY
	6/26/01	CFIA / Saskatoon Lab

FOREIGN GOV'T AGENCY	CITY & COUNTRY	ADDRESS OF LABORATORY
CFIA	Saskatoon, Canada	

NAME OF REVIEWER	NAME OF FOREIGN OFFICIAL
Dr. Oto Urban	Dr. James D. MacNeil

RESIDUE	ITEM	COMMENTS
IVE	3	Sometimes the results are analyzed in excess of 14 working dates time frame.
100, 200, 203, 300, 500, IVE	15	The lab is not providing the name of individual analysts performing the required intralaboratory check samples.